

How to clean the Dräger Caleo®



Please reprocess the device according to your hospital's guidelines. This document is for information purposes only. Please refer to the device instructions for use for complete disassembly, assembly, cleaning and surface disinfection. Only use disinfectants as listed in the instructions for use. Ensure that no liquid enters the device.

Prerequisites: There must be no patient in the device. Allow device to cool for 30 minutes prior to disassembly.

Surface disinfection for:

- Basic device/trolley
- · All disassembled parts
- Hood and side walls (interior and exterior)
- X-ray valve air channels: Pull a cloth soaked in disinfectant through each air channel

Machine cleaning and thermal disinfection:


- Grommets
- Fan impeller
- Water tank (option)


Before starting use, ensure that all surfaces are dry and free from disinfectant residues.

Disassembly ►














Assembly



12 Close side walls.

8 Remove grommets.



11 Insert grommets.

9 Open head and foot access ports.



10 Close head and foot access ports.

10 Remove mattress.



9 Insert mattress.

11 Remove mattress tray.



8 Insert mattress tray.

Remove hose holder arm (option).



7 Insert hose holder arm (option).

Unlock and remove x-ray tray.



6 Insert and lock x-ray tray.

14 Remove trough.



5 Insert trough.

Remove fan impeller and connecting piece.



4 Insert fan impeller and connecting piece.

16 Remove drawer.



3 Insert drawer.

17 Clean and disinfect basic device in full.



2

Remove display unit rotary knob and clean and disinfect display unit.



1 Insert display unit rotary knob.

Assembly